

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 04/22/2020

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:48
INTERNAL STANDARD	VERIFIED	00:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:49
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:50
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52

