

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 04/06/2020

OPERATOR NAME:  
CORBET W SULLIVAN  
Permit #: 45898  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:10
INTERNAL STANDARD	VERIFIED	15:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:10
BLANK TEST	0.00	15:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:13
BLANK TEST	0.00	15:14

CS  
#623

