

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 04/20/2020

OPERATOR NAME:
ETHAN J MACURA
Permit #: 44080
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 14:37 |
| INTERNAL STANDARD | VERIFIED | 14:37 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 14:37 |
| BLANK TEST | 0.00 | 14:38 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 14:39 |
| BLANK TEST | 0.00 | 14:40 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 14:40 |
| BLANK TEST | 0.00 | 14:41 |

#096
EJM

