

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 05/09/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:44
INTERNAL STANDARD	VERIFIED	13:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:45
BLANK TEST	0.00	13:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:46
BLANK TEST	0.00	13:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:47
BLANK TEST	0.00	13:48

