

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 05/13/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:45
INTERNAL STANDARD	VERIFIED	19:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:46
BLANK TEST	0.00	19:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:47
BLANK TEST	0.00	19:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:48
BLANK TEST	0.00	19:49

