

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 05/01/2020

OPERATOR NAME:  
CORY L HURLBURT  
Permit #: 37017  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:04
INTERNAL STANDARD	VERIFIED	09:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:06
BLANK TEST	0.00	09:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:08
BLANK TEST	0.00	09:09

CH # 257

