

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 05/04/2020

OPERATOR NAME:  
KYLE I KOLAR  
Permit #: 44063  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:17
INTERNAL STANDARD	VERIFIED	18:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:17
BLANK TEST	0.00	18:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21

*Handwritten signature and number #691*

