

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 04/27/2020

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:53
INTERNAL STANDARD	VERIFIED	14:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:53
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:54
BLANK TEST	0.00	14:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:55
BLANK TEST	0.00	14:56

