

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 05/11/2020

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:58
INTERNAL STANDARD	VERIFIED	05:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:58
BLANK TEST	0.00	05:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:59
BLANK TEST	0.00	06:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:00
BLANK TEST	0.00	06:01

