

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 04/13/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:36
INTERNAL STANDARD	VERIFIED	22:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:37
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:40

