

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 04/27/2020

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:41
INTERNAL STANDARD	VERIFIED	19:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:41
BLANK TEST	0.00	19:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:43
BLANK TEST	0.00	19:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:44
BLANK TEST	0.00	19:45

