

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 05/20/2020

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:05
INTERNAL STANDARD	VERIFIED	00:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:07
BLANK TEST	0.00	00:08

