

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 05/30/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:56
INTERNAL STANDARD	VERIFIED	05:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:58
BLANK TEST	0.00	05:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:59
BLANK TEST	0.00	06:00

