

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 06/17/2020

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:02
INTERNAL STANDARD	VERIFIED	00:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:05

