

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 06/13/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:50
INTERNAL STANDARD	VERIFIED	06:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:53
BLANK TEST	0.00	06:54

