

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 06/02/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:10
INTERNAL STANDARD	VERIFIED	20:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14

