

SUPERVISOR TEST

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 05/31/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:08
INTERNAL STANDARD	VERIFIED	19:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:10
BLANK TEST	0.00	19:11

