

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 06/29/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:11
INTERNAL STANDARD	VERIFIED	12:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:12
BLANK TEST	0.00	12:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:13
BLANK TEST	0.00	12:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:14
BLANK TEST	0.00	12:15

