

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 07/11/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:32
INTERNAL STANDARD	VERIFIED	17:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:32
BLANK TEST	0.00	17:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:34
BLANK TEST	0.00	17:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:35
BLANK TEST	0.00	17:36

