

SOLUTION CHANGE

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 07/05/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:43
INTERNAL STANDARD	VERIFIED	16:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:43
BLANK TEST	0.00	16:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:44
BLANK TEST	0.00	16:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:45
BLANK TEST	0.00	16:46

