

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 07/13/2020

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:34
INTERNAL STANDARD	VERIFIED	08:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:37
BLANK TEST	0.00	08:38

