

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 07/07/2020

OPERATOR NAME:
 BRAD M LYON
 Permit #: 44978
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:49
INTERNAL STANDARD	VERIFIED	18:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:50
BLANK TEST	0.00	18:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:52
BLANK TEST	0.00	18:53

