

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 07/14/2020

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:53
INTERNAL STANDARD	VERIFIED	20:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:53
BLANK TEST	0.00	20:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:55
BLANK TEST	0.00	20:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:56
BLANK TEST	0.00	20:57

