

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/07/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:18
INTERNAL STANDARD	VERIFIED	01:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:22

