

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 06/29/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:17
INTERNAL STANDARD	VERIFIED	20:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:17
BLANK TEST	0.00	20:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:19
BLANK TEST	0.00	20:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:20
BLANK TEST	0.00	20:21

