

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dnt: 106406

Date: 05/12/2020

OPERATOR NAME:  
TRAVIS A EARL  
Permit #: 33337  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:04
INTERNAL STANDARD	VERIFIED	12:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:04
BLANK TEST	0.00	12:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:06
BLANK TEST	0.00	12:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:07
BLANK TEST	0.00	12:08

*TAE*  
#68

