

SOLUTION CHANGE

-----  
 WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406  
 -----

Date: 07/01/2020

OPERATOR NAME:  
 KATHERINE M PADDOCK  
 Permit #: 45899  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:15
INTERNAL STANDARD	VERIFIED	20:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:15
BLANK TEST	0.00	20:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:17
BLANK TEST	0.00	20:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:18
BLANK TEST	0.00	20:19

*K.P. Paddock*

