

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/06/2020

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:54
INTERNAL STANDARD	VERIFIED	06:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:55
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:58
BLANK TEST	0.00	06:59

CS
#623

