

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 07/20/2020

OPERATOR NAME:  
KYLE L KOLAR  
Permit #: 44063  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:37
INTERNAL STANDARD	VERIFIED	21:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:38
BLANK TEST	0.00	21:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:39
BLANK TEST	0.00	21:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:41
BLANK TEST	0.00	21:42

*[Handwritten Signature]* #691

