

SUPERVISOR TEST

-----  
 WASHINGTON CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100284  
 -----

Date: 06/15/2020

OPERATOR NAME:  
 D TODD RICHE  
 Permit #: 33411  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:48
INTERNAL STANDARD	VERIFIED	07:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:48
BLANK TEST	0.00	07:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:49
BLANK TEST	0.00	07:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:50
BLANK TEST	0.00	07:51

