

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 06/22/2020

OPERATOR NAME:
 MATTHEW K JACKSON
 Permit #: 40078
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:39
INTERNAL STANDARD	VERIFIED	13:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:39
BLANK TEST	0.00	13:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:40
BLANK TEST	0.00	13:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:41
BLANK TEST	0.00	13:42

