

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 07/21/2020

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:20
INTERNAL STANDARD	VERIFIED	07:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23

