

SOLUTION CHANGE

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 08/01/2020

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:58
INTERNAL STANDARD	VERIFIED	06:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:58
BLANK TEST	0.00	06:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:00
BLANK TEST	0.00	07:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:01

