

```
<?xml version="1.0" standalone="yes"?>
<TRACS_IMPORT>
  <ALCOTEST_VALUES>
    <SERIAL>ARAD-0019</SERIAL>
    <IP_ADRESS>10.81.80.156</IP_ADRESS>
    <DATE>11/04/2019</DATE>
    <TIME>04:30PM</TIME>
  </ALCOTEST_VALUES>
  <CALCHECK_1>
    <CAL_MESSUNG_2_MESSUNG_ERROR>0</CAL_MESSUNG_2_MESSUNG_ERROR>
    <CAL_MESSUNG_2_IR_AAK_NULLWERT_1>0.00</CAL_MESSUNG_2_IR_AAK_NULLWERT_1>
    <CAL_MESSUNG_2_IR_AAK_MESSWERT>0.10</CAL_MESSUNG_2_IR_AAK_MESSWERT>
    <CAL_MESSUNG_2_EC_AAK_MESSWERT>0.09</CAL_MESSUNG_2_EC_AAK_MESSWERT>
    <CAL_MESSUNG_2_IR_AAK_NULLWERT_2>0.00</CAL_MESSUNG_2_IR_AAK_NULLWERT_2>
    <ID_BLANK_TIME_BEFORE_2>04:19:08PM 11/04/2019</ID_BLANK_TIME_BEFORE_2>
    <ID_TIME_START_BLOW_2>--:--:-- --/--/----</ID_TIME_START_BLOW_2>
    <ID_TIME_MEAS_2>04:19:45PM 11/04/2019</ID_TIME_MEAS_2>
    <ID_BLANK_TIME_AFTER_2>04:20:17PM 11/04/2019</ID_BLANK_TIME_AFTER_2>
  </CALCHECK_1>
  <SUBJECT_TEST>
    <PROB_MESSUNG_2_MESSUNG_ERROR>0</PROB_MESSUNG_2_MESSUNG_ERROR>
    <PROB_MESSUNG_2_IR_AAK_NULLWERT_1>0.00</PROB_MESSUNG_2_IR_AAK_NULLWERT_1>
    <PROB_MESSUNG_2_IR_AAK_MESSWERT>0.00</PROB_MESSUNG_2_IR_AAK_MESSWERT>
    <PROB_MESSUNG_2_EC_AAK_MESSWERT>0.00</PROB_MESSUNG_2_EC_AAK_MESSWERT>
    <PROB_MESSUNG_2_IR_AAK_NULLWERT_2>0.00</PROB_MESSUNG_2_IR_AAK_NULLWERT_2>
    <PROB_MESSUNG_2_BLAS_FEHLVERSUCHE>1</PROB_MESSUNG_2_BLAS_FEHLVERSUCHE>
    <PROB_MESSUNG_2_BLAS_VOLUMEN>1.9</PROB_MESSUNG_2_BLAS_VOLUMEN>
    <PROB_MESSUNG_2_BLAS_ZEIT>9.3</PROB_MESSUNG_2_BLAS_ZEIT>
    <ID_BLANK_TIME_BEFORE_3>04:25:20PM 11/04/2019,04:21:25PM 11/04/2019</ID_BLANK_TIME_BEFORE_3>
    <ID_TIME_START_BLOW_3>04:26:10PM 11/04/2019</ID_TIME_START_BLOW_3>
    <ID_TIME_END_BLOW_3>04:26:19PM 11/04/2019</ID_TIME_END_BLOW_3>
    <ID_TIME_MEAS_3>04:26:28PM 11/04/2019,04:24:38PM 11/04/2019</ID_TIME_MEAS_3>
    <ID_BLANK_TIME_AFTER_3>04:27:19PM 11/04/2019</ID_BLANK_TIME_AFTER_3>
  </SUBJECT_TEST>
  <CALCHECK_2>
    <CAL_MESSUNG_3_MESSUNG_ERROR>0</CAL_MESSUNG_3_MESSUNG_ERROR>
    <CAL_MESSUNG_3_IR_AAK_NULLWERT_1>0.00</CAL_MESSUNG_3_IR_AAK_NULLWERT_1>
    <CAL_MESSUNG_3_IR_AAK_MESSWERT>0.10</CAL_MESSUNG_3_IR_AAK_MESSWERT>
    <CAL_MESSUNG_3_EC_AAK_MESSWERT>0.09</CAL_MESSUNG_3_EC_AAK_MESSWERT>
    <CAL_MESSUNG_3_IR_AAK_NULLWERT_2>0.00</CAL_MESSUNG_3_IR_AAK_NULLWERT_2>
    <ID_BLANK_TIME_BEFORE_4>04:27:19PM 11/04/2019</ID_BLANK_TIME_BEFORE_4>
    <ID_TIME_START_BLOW_4>--:--:-- --/--/----</ID_TIME_START_BLOW_4>
    <ID_TIME_MEAS_4>04:27:21PM 11/04/2019</ID_TIME_MEAS_4>
    <ID_BLANK_TIME_AFTER_4>04:27:52PM 11/04/2019</ID_BLANK_TIME_AFTER_4>
  </CALCHECK_2>
  <GENERAL_DATA>
    <ID_SERIAL_NO>ARAD-0019</ID_SERIAL_NO>
    <ID_SW_VERSION> 8321288 1.12.2</ID_SW_VERSION>
    <ID_START_TIME>04:14:05PM 11/04/2019</ID_START_TIME>
    <ID_END_TIME>04:28:47PM 11/04/2019</ID_END_TIME>
    <ID_LAST_CAL_DATE>04/23/2019</ID_LAST_CAL_DATE>
    <ID_AAK_ENDRESULT> 0.00 %BAC</ID_AAK_ENDRESULT>
    <ID_TEST_COUNTER>743</ID_TEST_COUNTER>
    <ID_MEAS_PROC>1</ID_MEAS_PROC>
    <ID_CALIBRATION_COUNTER>93</ID_CALIBRATION_COUNTER>
    <ID_ERROR_PROTOKOLL>0</ID_ERROR_PROTOKOLL>
    <ID_TEST_GAS_TYPE>2</ID_TEST_GAS_TYPE>
```

<ID_TEST_GAS_INLET>1</ID_TEST_GAS_INLET>
<ID_TEST_GAS_CONC>25.00</ID_TEST_GAS_CONC>
<ID_TEST_CFG_SENSOR>0</ID_TEST_CFG_SENSOR>
<ID_DATA_INPUT_ON_OFF>1</ID_DATA_INPUT_ON_OFF>
<ID_IR_AVERAGE>21227</ID_IR_AVERAGE>
<ID_EC_SENSOR_OFFSET>109</ID_EC_SENSOR_OFFSET>
<ID_HOSE_TEMP>47.0</ID_HOSE_TEMP>
<ID_IR_SENSOR_TEMP>45.3</ID_IR_SENSOR_TEMP>
<ID_INTERNAL_DEVICE_TEMP>36.6</ID_INTERNAL_DEVICE_TEMP>
<ID_12V_VOLTAGE>12.1</ID_12V_VOLTAGE>
<ID_3V_BACKUP_VOLTAGE>3.3</ID_3V_BACKUP_VOLTAGE>
<ID_MS_SW_CRC>0x6581</ID_MS_SW_CRC>
<ID_PRESSURE_DRY_GAS>0</ID_PRESSURE_DRY_GAS>
<ID_DRY_GAS_LOT_NO>979115 CYL 77</ID_DRY_GAS_LOT_NO>
<ID_DRY_GAS_EXP_DATA>07/05/2021</ID_DRY_GAS_EXP_DATA>
</GENERAL_DATA>
<INPUT_VALUES>
<UF_CASE_NUMBER>9246706</UF_CASE_NUMBER>
<ZERO_TOLERANCE_CASE>NO</ZERO_TOLERANCE_CASE>
<TROOP>G</TROOP>
<STATION>Granville</STATION>
<ARREST_DATE>11/04/2019</ARREST_DATE>
<ARREST_TIME>03:30PM</ARREST_TIME>
<ARRESTING_MEM_LAST>Hatalsky</ARRESTING_MEM_LAST>
<ARRESTING_MEM_FIRST>Nathan</ARRESTING_MEM_FIRST>
<ARRESTING_MEM_MI>M</ARRESTING_MEM_MI>
<ARRESTING_MEM_SHIELD>3274</ARRESTING_MEM_SHIELD>
<SUBJECT_LAST_NAME>Lambert</SUBJECT_LAST_NAME>
<SUBJECT_FIRST_NAME>Mae</SUBJECT_FIRST_NAME>
<SUBJECT_MI>A</SUBJECT_MI>
<SUBJECT_DOB>09/29/1970</SUBJECT_DOB>
<SUBJECT_AGE>49</SUBJECT_AGE>
<SUBJECT_GENDER>FEMALE</SUBJECT_GENDER>
<OPERATOR_LAST_NAME>Hatalsky</OPERATOR_LAST_NAME>
<OPERATOR_FIRST_NAME>Nathan</OPERATOR_FIRST_NAME>
<OPERATOR_MI>M</OPERATOR_MI>
<OPERATOR_SHIELD_NO>3274</OPERATOR_SHIELD_NO>
<OPERATOR_PERMIT_NO>47897</OPERATOR_PERMIT_NO>
<OPERATOR_PERMIT_EXP>06/14/2021</OPERATOR_PERMIT_EXP>
<SUBJECT_OBSERVED_20_MINUTES>YES</SUBJECT_OBSERVED_20_MINUTES>
<TEST_20_MINUTES_PAST_ARREST>YES</TEST_20_MINUTES_PAST_ARREST>
<REFUSAL_MIRANDA_READ>YES</REFUSAL_MIRANDA_READ>
</INPUT_VALUES>
</TRACS_IMPORT>