

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/11/2020

OPERATOR NAME:
RYAAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:32
INTERNAL STANDARD	VERIFIED	03:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:34
BLANK TEST	0.00	03:35

