

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/05/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:09
INTERNAL STANDARD	VERIFIED	19:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:11
BLANK TEST	0.00	19:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:12
BLANK TEST	0.00	19:12

