

SOLUTION CHANGE

-----  
CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706  
-----

Date: 10/05/2020

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:07
INTERNAL STANDARD	VERIFIED	08:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:09
BLANK TEST	0.00	08:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:10
BLANK TEST	0.00	08:11

