

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 09/28/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:59
INTERNAL STANDARD	VERIFIED	08:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:59
BLANK TEST	0.00	09:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:00
BLANK TEST	0.00	09:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:01
BLANK TEST	0.00	09:02

