

SUPERVISOR TEST

-----  
 FORT EDWARD  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 115708  
 -----

Date: 09/15/2020

OPERATOR NAME:  
 SGT DEAN E WATKINS  
 Permit #: 39523  
 Operator Agency/Dept: FORT EDWARD VILL. PD  
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:51
INTERNAL STANDARD	VERIFIED	09:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:53
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:54
BLANK TEST	0.00	09:55

