

SOLUTION CHANGE

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 08/04/2020

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:21
INTERNAL STANDARD	VERIFIED	19:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:24
BLANK TEST	0.00	19:25

