

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/11/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:48
INTERNAL STANDARD	VERIFIED	18:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:50
BLANK TEST	0.00	18:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:51
BLANK TEST	0.00	18:52

