

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 10/28/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:39
INTERNAL STANDARD	VERIFIED	18:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:40
BLANK TEST	0.00	18:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:41
BLANK TEST	0.00	18:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:42
BLANK TEST	0.00	18:43

