

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 10/16/2020

OPERATOR NAME:
BRAD M LYON
Permit #: 44978
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:29
INTERNAL STANDARD	VERIFIED	07:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:30
BLANK TEST	0.00	07:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:31
BLANK TEST	0.00	07:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:32
BLANK TEST	0.00	07:33

