

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 10/07/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20340\

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:01
INTERNAL STANDARD	VERIFIED	05:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:01
BLANK TEST	0.00	05:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:02
BLANK TEST	0.00	05:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:04
BLANK TEST	0.00	05:04

