

SOLUTION CHANGE

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 09/30/2020

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:18
INTERNAL STANDARD	VERIFIED	18:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:22

