

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 09/16/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:51
INTERNAL STANDARD	VERIFIED	18:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:51
BLANK TEST	0.00	18:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:52
BLANK TEST	0.00	18:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:54
BLANK TEST	0.00	18:55

