

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 09/26/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:05
BLANK TEST	0.00	07:05

