

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 09/19/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:34
INTERNAL STANDARD	VERIFIED	10:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:37
BLANK TEST	0.00	10:37

