

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 09/06/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:49
INTERNAL STANDARD	VERIFIED	07:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:50
BLANK TEST	0.00	07:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:51
BLANK TEST	0.00	07:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:52
BLANK TEST	0.00	07:53

