

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/28/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:11
INTERNAL STANDARD	VERIFIED	20:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14

